

Oro Clean Chemie AG Allmendstrasse 21 8320 Fehraltorf

Efficacy of DENTIRO® Sensitive against Candida albicans ATCC 10231 in a quantitative suspension test at 20 °C according to EN 13624:2013 (E) under dirty condition

## **EXPERT OPINION**

This expert opinion is based on the test report VX-TR-18-0300 dated 16 March 2018.

The yeasticidal activity of the disinfectant DENTIRO® Sensitive of Oro Clean Chemie AG against Candida albicans ATCC 10231 was investigated by a quantitative suspension test according to EN 13624:2013 (E) under dirty condition (3.00 g/L bovine albumin solution and 3.00 ml/L sheep erythrocytes).

According to this suspension test, a disinfectant or a disinfectant solution at a particular concentration is considered as having yeasticidal activity if the number of viable yeast is reduced by ≥4 log₁₀ (inactivation ≥99.99%) within the recommended exposure period.

DENTIRO® Sensitive was examined at 20 °C at the concentration of 100.00 %\* for the exposure time of 15 seconds. After the exposure time, the yeasticidal reduction exceeded 4 log10-steps in all assays. Therefore, a yeasticidal activity against Candida albicans ATCC 10231 was measured as follows:

> 100.00 %\* 15 seconds Dirty condition

Kuala Lumpur, 16 March 2018

There

**Dr Cheng Shoou Lee** 

Team Leader Bacteriology Testing

**Zhao Min Khoo** Microbiologist

<sup>\*</sup> The product can only be tested at 80.00 % concentration or less, as some dilution always occurs when test organisms and interfering substance are added.